
 Government eProcurement System		Government eProcurement System Published Corrigendum Details			
					Date : 03-Mar-2021 03:45 PM
					Print
Organisation Chain :	Council of Scientific and Industrial Research NEIST-Jorhat - CSIR Purchase-NEIST - CSIR				
Tender ID :	2021_CSIR_70538_1				
Tender Ref No :	3(POQ)/108/20-21/PUR				
Tender Title :	Supply and Installation of UV-Visible Spectrophotometer at CSIR NEIST Branch Laboratory, Itanagar				
Corrigendum Type :	Technical Bid				
Corrigendum Document Details					
Corr.No.	Corrigendum Title	Corrigendum Description	Published Date	Document Name	Doc Size(in KB)
1	Technical Corrigendum	Technical Corrigendum	03-Mar-2021 03:45 PM	Corrigendum.pdf 	262.48

**CSIR - North East Institute of Science and Technology
Jorhat, Assam**

This is to inform that in the Tender for "Supply and Installation of UV Visible Spectrophotometer" having **Tender ID "2021_CSIR_70538_1"** the following have been amended-

In Chapter – 3, Point No. B. Specifications and Allied Technical Details -

Earlier Specifications -	Is amended as -
Detectors: Silicon Diodes	Detector: Silicon Diodes/Photomultiplier (PMT)
Monochromator: Holographic/diffraction grating or Double out of plane Littrow monochromator	Monochromator: Holographic / diffraction grating on Czerny Turner monochromator or Littrow monochromator
2 Pack of 3.5 ml quartz cell with stopper	2 pair of 3.5 ml quartz cell with stopper
Temperature Controlled System, Water free temperature from 5°C to 80°C or better	Temperature Controlled System: Water/Water free temperature cycling from 7°C to 80°C or better

All other specs and terms & conditions of the NIT will remain same